



DIVISION OF INDUSTRY SERVICES  
Plumbing Product Review  
P.O. Box 2658  
Madison, Wisconsin 53701-2658  
**TTY: Contact Through Relay**

**Scott Walker, Governor**  
**Dave Ross, Secretary**

February 15, 2013

STERIS CORPORATION  
CHAD BITTNER  
5960 HEISLEY ROAD  
MENTOR OH 44060

Re: Description: HEALTH CARE PLUMBING APPLIANCE  
Manufacturer: STERIS CORPORATION  
Product Name: AMSCO STEAM STERILIZERS  
Model Number(s): CENTURY SMALL (16 & 20"), MILLENNIUM SMALL (16 & 20"), 400 SMALL (16 & 20"),  
C SMALL (16 & 20"), CENTURY MEDIUM (36, 39, 48, 49, 60 & 61"), 400 MEDIUM  
(36, 39, 48, 49, 60 & 61"), EVOLUTION MEDIUM (39, 42, 49, 54, 61 & 66") AND  
EVOLUTION FLOOR LOADER (49 & 72").  
Product File No: 20120404

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of February 2018.

This approval is contingent upon compliance with the following stipulation(s):

- A reduced pressure principle backflow preventer (ASSE 1013 or CAN/CSA B64.4) or air-gap must be installed on the water supply pipes serving this product.
- The indirect waste pipe from this product must discharge by means of an air-gap, an air-gap fitting that conforms to ASME A112.1.3 or an air-gap fitting approved by the Department.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter  
Plumbing Product Reviewer  
Department of Safety and Professional Services  
Division of Industry Services  
Bureau of Technical Services  
(608) 267-1401 Phone  
(608) 266-2602 Fax  
glen.schlueter@wi.gov E-mail